

Fusedrawer für Fuse Links 5x20 mm, with or without Voltage Selector Insert



Fuse drawer without voltage selector insert with shorting bar (Fuse links not enclosed)

Fuse drawer with voltage selector insert (Fuse links not enclosed)

### Approvals and Compliances

#### Description

- Component
- Snap-in version from front side
- 2 Functions :
- Fuseholder for Fuse links 5x20 mm , Voltage Selector
- Approved together with power entry module

#### Characteristics

- Fuse drawer meets requirements of medical standard IEC/EN 60601-1
- Suitable for use in equipment according to IEC/UL 60950

#### Other versions on request

- Fuse drawer with other voltage values

#### References

Fits to type: [DD11](#); [DD12](#); [DD14](#); [DD21](#); [DD22](#); [KEA-Print](#); [KFA](#); [KMF](#)

#### Weblinks

[pdf datasheet](#), [html-datasheet](#), [General Product Information](#), [Distributor Stock-Check](#), [Accessories](#), [Detailed request for product](#), [Microsite](#)

### Approvals and Compliances

Detailed information on product approvals, code requirements, usage instructions and detailed test conditions can be looked up in [Details about Approvals](#)

#### Application standards

Application standards where the product can be used

| Organization | Design                         | Standard     | Description   |
|--------------|--------------------------------|--------------|---|
|              | Designed for applications acc. | IEC/UL 60950 | IEC 60950-1 includes the basic requirements for the safety of information technology equipment.   |
|              | Designed for applications acc. | IEC 60601-1  | Medical electrical equipment - Part 1: General requirements for basic safety and essential performance  |
|              | Designed for applications acc. | IEC 60335-1  | Safety of electrical appliances for household and similar purposes. Meets the requirements for appliances in unattended use. This includes the enhanced requirements of glow wire tests acc. to IEC 60695-2-12 and -13. |

#### Compliances

The product complies with following Guide Lines

| Identification | Details           | Initiator   | Description   |
|----------------|-------------------|-------------|---|
|                | RoHS              | SCHURTER AG | EU Directive RoHS 2011/65/EU  |
|                | China RoHS        | SCHURTER AG | The law SJ / T 11363-2006 (China RoHS) has been in force since 1 March 2007. It is similar to the EU directive RoHS.  |
|                | REACH             | SCHURTER AG | On 1 June 2007, Regulation (EC) No 1907/2006 on the Registration, Evaluation, Authorization and Restriction of Chemicals 1 (abbreviated as "REACH") entered into force. |
|                | Medical Equipment | SCHURTER AG | Suitable for use in medical equipment according to IEC/UL 60601-1   |

## Voltage Selector Inserts

| xx           | Marking on Voltage Selector Insert / Terminal No. |     |     |     |
|--------------|---|-----|-----|-----|
|              | 1   | 2   | 3   | 4   |
| 4301.1014.01 | 110   | 150 | 220 | -   |
| 4301.1014.02 | 120   | -   | 240 | -   |
| 4301.1014.03 | 110   | -   | 220 | -   |
| 4301.1014.04 | 115   | -   | 220 | -   |
| 4301.1014.05 | 110   | -   | 230 | -   |
| 4301.1014.06 | 115   | -   | 230 | -   |
| 4301.1014.07 | 100   | 110 | 220 | 240 |
| 4301.1014.08 | 100   | 120 | 220 | 240 |
| 4301.1014.10 | 220   | -   | 240 | -   |
| 4301.1014.11 | 120   | 220 | 240 | -   |
| 4301.1014.12 | 110   | 220 | 240 | -   |
| 4301.1014.14 | -   | -   | -   | -   |
| 4301.1014.15 | 100   | 120 | 220 | -   |
| 4301.1014.23 | 100   | 115 | 220 | 240 |
| 4301.1014.24 | 115   | 230 | 240 | -   |
| 4301.1014.29 | 100   | 120 | 230 | 240 |
| 4301.1014.34 | 100   | 115 | 230 | 240 |
| 4301.1014.49 | 110   | 220 | 230 | 240 |
| 4301.1014.51 | 100   | -   | -   | -   |
| 4301.1014.52 | 120   | 220 | 230 | 240 |
| 4301.1014.53 | 100   | -   | -   | -   |
| 4301.1014.54 | 120   | 220 | 230 | 240 |
| 4301.1024.02 | 120   | -   | 240 | -   |
| 4301.1024.03 | 110   | -   | 220 | -   |
| 4301.1024.05 | 110   | -   | 230 | -   |
| 4301.1024.06 | 115   | -   | 230 | -   |
| 4301.1024.07 | 100   | 110 | 220 | 240 |
| 4301.1024.08 | 100   | 120 | 220 | 240 |
| 4301.1024.11 | 120   | 220 | 240 | -   |
| 4301.1024.12 | 110   | 220 | 240 | -   |
| 4301.1024.14 | -   | -   | -   | -   |
| 4301.1024.15 | 100   | 120 | 220 | -   |
| 4301.1024.23 | 100   | 115 | 220 | 240 |
| 4301.1024.24 | 115   | 230 | 240 | -   |
| 4301.1024.25 | 100   | 120 | 230 | -   |
| 4301.1024.26 | 220   | 230 | 240 | -   |
| 4301.1024.36 | 100   | 110 | 120 | -   |
| 4301.1024.37 | 100   | 110 | 230 | 240 |
| 4301.1024.38 | 110   | 127 | 230 | -   |
| 4301.1024.45 | 100   | 220 | 230 | 240 |
| 4301.1024.47 | 100   | -   | 120 | -   |
| 4301.1024.48 | 100   | 110 | 230 | -   |
| 4301.1214.01 | 110   | 150 | 220 | -   |
| 4301.1214.02 | 120   | -   | 240 | -   |
| 4301.1214.03 | 110   | -   | 220 | -   |
| 4301.1214.05 | 110   | -   | 230 | -   |
| 4301.1214.06 | 115   | -   | 230 | -   |
| 4301.1214.07 | 100   | 110 | 220 | 240 |
| 4301.1214.08 | 100   | 120 | 220 | 240 |
| 4301.1214.10 | 220   | -   | 240 | -   |
| 4301.1214.11 | 120   | 220 | 240 | -   |
| 4301.1214.12 | 110   | 220 | 240 | -   |
| 4301.1214.13 | 115   | 220 | 240 | -   |
| 4301.1214.14 | -   | -   | -   | -   |
| 4301.1214.15 | 100   | 120 | 220 | -   |